



SAFETY AND BUILDINGS DIVISION
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Jim Doyle, Governor
Richard J. Leinenkugel, Secretary

December 30, 2008

A-DEC INC.
ENGINEERING
TOM LOUISELL
2601 CRESTVIEW DRIVE
PO BOX 111
NEWBERG OR 97132

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: A-DEC INC.
Product Name: VAC BACK CHAIR
Model Number(s): DECADE VAC BACK CHAIR, 61-1437-00, PRIORITY VAC BACK CHAIR, 61-0514-01
AND VAC BACK CHAIR 1040, 1020, 1020 SOLO AND 1021
Product File No: 20080372

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters Comm 82 through 84, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of May 2014.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or air-gap must be installed on the water supply pipes serving this product.

This approval supersedes the approval issued on May 4, 2004 under product file number 20040193.

This approval letter shall be incorporated with your previously approved plans and/or specifications approved under product file number 20040193.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson
Plumbing Product Reviewer
Phone: 608-266-6742
Fax: 608-267-9566
E-mail: Jerry.Thompson@wi.gov